



Republic of the Philippines
DEPARTMENT OF AGRICULTURE WESTERN VISAYAS
Hamungaya, Brgy. Buntatala, Jaro, Iloilo City
westernvisayas@mail.da.gov.ph | (033) 336-4221

SUPPLEMENTAL/BID BULLETIN

This Addendum is issued to modify or amend items in the Bid Documents (Invitation to Bid for the Procurement of Supply and Delivery of Liquid Nitrogen (LN2) Field Tank – IB No. 2024-009 (GAS))

On Invitation to Bid: Specifications

- Maximum Storage Capacity
 - No. of Canisters: 6
 - Straw Holding Capacity Single Level (0.5ml): 540-879
- Performance
 - LN2 Capacity (Liter): 3.5 – 3.6
 - Static Evaporation Rate (L/day): 0.12 – 0.15
 - Static Holding Time (Days): 23 – 25
 - Normal Working Duration (Days): 14 – 15
- Unit Dimensions
 - Neck Opening (mm): 51 – 55
 - Overall Height (mm): 390 – 460
 - Outer Diameter (mm): 220 – 224
 - Canister Height (mm): 125 – 127
 - Canister Diameter (mm): 39 – 41

Terms and Conditions:

- One (1) year warranty (in case of leakage, supplier should replace the field tank)
- With quality waterproof sticker of DA-Western Visayas Logo (size: 12cm x 12cm)

On Eligibility and Technical Requirements

7.1 Copy of the license of the Surety Firm in case a surety bond is filed

8. Audited Financial Statement

- 8.1** Auditors Opinion,
- 8.2** Income Statement,
- 8.3** Balance Sheet and
- 8.4** Cash Flow);



Republic of the Philippines
DEPARTMENT OF AGRICULTURE WESTERN VISAYAS
 Hamungaya, Brgy. Buntatala, Jaro, Iloilo City
 westernvisayas@mail.da.gov.ph | (033) 336-4221

On Philippine Bidding Documents: Section VII. Technical Specifications

	27 Tanks Liquid Nitrogen (LN2) Field Tank	Comply
	<ul style="list-style-type: none"> • Maximum Storage Capacity <ul style="list-style-type: none"> ➤ No. of Canisters: 6 ➤ Straw Holding Capacity Single Level (0.5ml): 540-879 • Performance <ul style="list-style-type: none"> ➤ LN2 Capacity (Liter): 3.5 – 3.6 ➤ Static Evaporation Rate (L/day): 0.12 – 0.15 ➤ Static Holding Time (Days): 23 – 25 ➤ Normal Working Duration (Days): 14 – 15 • Unit Dimensions <ul style="list-style-type: none"> ➤ Neck Opening (mm): 51 – 55 ➤ Overall Height (mm): 390 – 460 ➤ Outer Diameter (mm): 220 – 224 ➤ Canister Height (mm): 125 – 127 ➤ Canister Diameter (mm): 39 – 41 	Comply
	<p>Terms and Conditions:</p> <ul style="list-style-type: none"> • One (1) year warranty (in case of leakage, supplier should replace the field tank) • With quality waterproof sticker of DA-Western Visayas Logo (size: 12cm x 12cm) 	Comply

ZARLINA B. CUELLO
 BAC Chairperson